



Infusion solution stand for infusion therapy with manual height-adjustment.

IQI
57

TQI
14



- Indicated for Inpatient Wards, Outpatient Dept., Patient's Home
- Maximum height of the infusion support: 2200 mm
- Maximum capacity of the device: 4 Kg
- Infusions support with 4 safety hooks in **high visibility** Nylon
- Vertical telescopic structure in **Stainless** Anodised Aluminium
- Base with 5 spokes with **small clearance** in **Stainless** Anodised Aluminium

SA.F.E. Tech.

Effectiveness
PQI 47,2

Sanitizability
PQI 56,7

Usability
PQI 60,5


Versatility
PQI 50,0



Systemic Compatibility
PQI 59,8

Environ. Compatibility
PQI 100



**NEAR
BED**



Properties of the Device	Ref. Technical-Functional Feature	Features of the Device							Qif
Versatility	6.2.1	Destination environment	Home	Outpatient Dept.	Ward	N.M.R.	Intensive Care Unit	Operating Theatre	50,00
			OK	OK	OK				
Effectiveness	6.1.2	Maximum height of the Infusion Solution Support (Hsi Max) (+50 / -0)	2200			mm			77,74
Effectiveness	6.1.1	Maximum Capacity (Infusion Solution) (Kg)	Device Max Capacity (M)	4	Individual hook capacity		1	16,67	
Systemic Compatibility	6.5.5	Infusion Pumps (Accepted weight)	up to 1 Kg (Structural Rod)						10,00
Usability/Sanitizability	6.3.1 6.4.1	Height Adjustment of the Infusion Solution Support	Manual, with friction clamp and safety screw						60,00
Sanitizability	6.3.1	Infusion Solution Support	Nylon 6.6						70,00
Usability	6.4.2		N°4 accidental anti-fall hooks for the Solutions						100,00
Usability	6.4.2		Individual hook mechanical strength (Kg)				4		50,00
Usability	6.4.2		High visibility of the infusions support hooks (anti-contusion)						70,00
Usability	6.4.5		Ergonomics-Infusion Solutions Hook (Niosh Method)	(Hsi Min) Infusion Solution load height = 1300 mm from the ground					
Sanitizability	6.3.2	Vertical telescopic structure (Materials / Tech. / Dimensions)	Stainless Anodised Aluminium(Ø28 / Ø20) Clamp SA.F.E. Technology						80,00
Sanitizability	6.3.2	Ergonomics-Base (Tripping)	Stainless Anodised Aluminium Fitting SA.F.E. Technology						70,00
Usability	6.4.5		N°5 spokes (Smaller clearance)						75,00
Usability	6.4.4	Manageability of the Device (at Max Capacity)	N°5 twin wheels Ø 50 mm			High visibility (anti-tripping)			77,41
Sanitizability	6.3.3	Constraints to the Ground							0,00
Sanitizability	6.3.4	Accepted sanitization procedure	Manual/Atomization/Steam						60,00
Systemic Compatibility	6.5.1	Dimensions and geometries [mm] Weight [Kg]	Ø 486 x 1348/2200H			2,5 Kg			79,01
Usability	6.4.5	Stability index (Rst Max) <small>(force to apply to the base of the device to determine its overturning)</small>	1,86			Kg			11,74
Usability	6.4.3	Device for the slowed descent of the Infusion Solution Support	No						0,00
Systemic Compatibility	6.5.2	Devices for elimination of electrostatic charges/electromagnetic comp.	Applied						100,00
Systemic Compatibility	6.5.3	Insensitivity to magnetic fields by NMR	NO						40,00
Systemic Compatibility	6.5.4	Colours Available, Identification, Design	Option available						80,00
Environmental Compatibility	6.6	Resistance to fire/ Latex free / Recyclable	100%	100%	100%			100,00	
Systemic Compatibility	6.5.6	Accessories available in the range HID-ACS.....	0-i-1-2-7-SQ-KD						50,00
Duration through time	6.7	Warranty Option (see confirmation on economic offer)	2			Years			
		Medical device	93/42 CEE Conformity / Class 1 						

Property Quality Indicators (PQI)

	6.1		Effectiveness						47,2
	6.3		Sanitizability						56,7
	6.4		Usability						60,5
	6.2		Versatility						50,0
	6.5		Systemic Compatibility						59,8
	6.6		Environmental Compatibility						100,0
	$(5 \times (\text{Effectiveness} + \text{Sanitizability} + \text{Usability}) + 1 \times (\text{Versatility} + \text{Syst.Comp.} + \text{Environmental Comp.})) / 18$		Intrinsic Quality Indicator (IQI)						57
	6.7	TOTAL INTRINSIC QUALITY through TIME Indicator (TQI)						14	